



Two decades of water vapor measurements with the FISH fluorescence hygrometer: A review

Christian Rolf (1), Jessica Meyer (1,12), Cornelius Schiller (+), Susanne Rohs (1,2), Nicole Spelten (1), Armin Afchine (1), Martin Zöger (3), Nikolay Sitnikov (4), Troy D. Thornberry (5,6), Andrew W. Rollins (5,6), Zoltan Bozóki (7,8), David Tátrai (7,8), Volker Ebert (9,10), Benjamin Kühnreich (9,10), Peter Mackrodt (9), Ottmar Möhler (11), Harald Saathoff (11), Karen H. Rosenlof (5), and Martina Krämer (1)

(1) Forschungszentrum Jülich, IEK-7, Jülich, Germany (c.rolf@fz-juelich.de), (2) IEK 8, Forschungszentrum Jülich, 52425 Jülich, Germany, (3) DLR, FX, 82234 Oberpfaffenhofen, Germany, (4) Central Aerological Observatory, Dolgoprudny, Russia, (5) NOAA ESRL Chemical Sciences Division, Boulder, USA, (6) CIRES, University of Colorado Boulder, Boulder, CO, USA, (7) University of Szeged, Szeged, Hungary, (8) MTA-SZTE Research Group on Photoacoustic Spectroscopy, Szeged, Hungary, (9) Physikalisch-Technische Bundesanstalt (PTB), 38116 Braunschweig, Germany, (10) Technische Universität Darmstadt, 64287 Darmstadt, Germany, (11) IMK-AAF, Karlsruhe Institute of Technology, 76344 Eggenstein-Leopoldshafen, Germany, (12) now at: Institute of Energy and Environmental Technology (IUTA), 47229 Duisburg, Germany, (+) deceased

The Fast In-situ Stratospheric Hygrometer (FISH) is an airborne Lyman- α photofragment fluorescence hygrometer for accurate and precise measurement of total water mixing ratios (WMR) (gas phase + evaporated ice) in the upper troposphere and lower stratosphere (UT/LS) since almost two decades. Here, we present a comprehensive review of the measurement technique, calibration procedure, accuracy and reliability of FISH.

A crucial part for the FISH measurement quality is the regular calibration to a water vapor reference, namely the commercial frostpoint hygrometer DP30. In the frame of this work this frostpoint hygrometer is compared to German and British traceable metrological water standards and its accuracy is found to be 2-4 %. Overall, in the range from 4-1000 ppmv, the total accuracy of FISH was found to be 6-8 % as stated also in previous publications. For lower mixing ratios down to 1 ppmv, the uncertainty reaches a lower limit of 0.3 ppmv. For specific, non-atmospheric conditions, as set in experiments at the AIDA chamber -namely mixing ratios below 10 ppmv and above 100 ppmv in combination with high and low pressure conditions- the need to apply a modified FISH calibration evaluation has been identified. The new evaluation improves the agreement of FISH with other hygrometers to $\pm 10\%$ accuracy in the respective mixing ratio ranges. Further, a quality check procedure for high total water measurements in cirrus clouds at high pressures (400-500 hPa) is introduced. The performance of FISH in the field is assessed by reviewing intercomparisons of FISH water vapor data with other in-situ and remote sensing hygrometers over the last two decades. We find that the agreement of FISH with the other hygrometers has improved over that time span from overall up to $\pm 30\%$ or more to about $\pm 5-20\%$ @ < 10 ppmv and to $\pm 0-15\%$ @ > 10 ppmv.

We will show here that the robust and continuous calibration and operation procedures of the FISH instrument over the last two decades establish FISH as one of the core instruments for in-situ observations of water vapor in the UT/LS.